



# **Radioactive Material Identification**









## Objective



#### Provide emergency responders advanced training on the ORTEC Detective to acquire high quality spectral data for radioactive material identification

Note: DOE will provide 2-3 Detectives for the course. Each Detective control display will be projected on a large screen to ensure maximum observation and participation by all students. If the host Competent Authority has access to additional Detectives, it would benefit having them in class enabling more hands-on training opportunities for the students.



## **Reason for Identification**



#### Why conduct radionuclide identification?

- To identify the radioisotope(s) causing the alarm
- To aid in law enforcement investigation and alarm adjudication
- To determine if cargo is consistent with the manifest
- To provide responders better understanding of potential hazard
- To allow responders to take appropriate protective actions
- To aid in law enforcement forensics investigations
- To aid in planning source recovery operations




### **Radioisotope Identification**



High Resolution vs Low Resolution Gamma Spectroscopy

#### Resolution - "ability to resolve adjacent gamma peaks"



Nal

High Purity Germanium (HPGe) is the *Gold Standard* for radioisotope identification



### **Detective-EX**



#### **ORTEC<sup>™</sup>** Detective-EX Specifications


Gamma – HPGe crystal (15% efficiency)<sup>1</sup> - 5 cm diameter x 3 cm high Gamma - Geiger-Mueller tube Neutron - four <sup>3</sup>He tubes

- 10 cm x 1.2 cm diameter; 20 atm pressure
- Embedded in high density polyethylene

Multi-Channel Analyzer (MCA) – 8k channels Cooler - low power Stirling type Temperature - less than -245 F (-154 C) Cool down time - 12 hrs Internal 12 volt battery - 3 hrs duration

Size - 37 cm x 18 cm x 34 cm Weight - 11.75 kg






<sup>1</sup> Industry standard is to report the HPGe detection efficiency as compared to a 7.6 cm x 7.6 cm sodium iodide detector for the <sup>60</sup>Co 1332 keV gamma at 25 cm.



### **High Purity Germanium (HPGe)**



### Schematic of a typical detector






### **Electronics**



#### Signal processing steps



producing ionizing charge

a fast negative millivolt analog signal


given positive unipolar shape and 2-6 microsecond width

on pulse height (i.e. voltage) and displayed as a gamma spectrum



## Main Window/Advanced Setup







8



## **Operating Status Windows**

PocketPC



<u>File Z</u> oom <u>T</u> ools <u>H</u> elp	
Adv. 09/0	9/2011 11:57:29 PM
Setup S	torage Space: 12192
Live Time	0:00:56
Real Time	0:00:56
Dead Time	0.1%
Bias Voltage	-2498
+12V	12.00
-12V	-12.00
+3.3V	3.30
Detector Temp	122.70
Body Temp	33.98
Cold Tip Temp	115.00
Cooler Drive	6.80
Ion Pump	3305.13.0
Cooler Runtime	e 1023.99
	Back

System Status

**System Parameters** 



### Calibration



### **Three Step Process**


PocketPC     Ele Zoom Iools Help	PocketPC     File Zoom Tools Help	Elle Zoom Iools Help
<i>Adv.</i> 09/10/2011 12:33:02 AM <i>Setup</i> Battery Time: 164 min	Adv. 09/10/2011 12:33:26 AM Setup 🔴 Battery Time: 164 min	<i>Adv.</i> 09/10/2011 12:34:11 AM <i>Setup</i> Battery Time: 164 min
<b>1</b> Place Cs-137 in front of the Detective-EX	<b>1</b> Place Cs-137 in front of the Detective-EX	<b>1</b> Place Cs-137 in front of the Detective-EX
2. Start Calibration Now	2. Abort Calibration	2. Start Calibration Now
	Calibration in progress	Centroid: 661.73 keV
	Time remaining: 50 seconds	FWHM: 1.82 keV
		Cal. adjustment: -0.02%
		The calibration can be improved.
Gamma Count Rate = 1578 cps	Gamma Count Rate = 1686 cps	3. Use Improved Calibration
Restore Default Back	Restore Default Back	Restore Default Back
Step 1	Step 2	<b>Step 3</b> 10



### Modes



### Search, ID and SNM Modes





### **Advanced Setup Screens**





PocketPC     Image: P	PocketPC     Eile Zoom Iools Help	
<i>Adv.</i> 09/09/2011 11:59:49 PM <i>Setup</i> Battery Time: 164 min	Adv. 09/09/2011 1 Setup Storage Sp	1:54:19 PM pace: 12192
y Dose Rate: Threshold: 10 Set Units: () µSv/hr () mrem/hr	Pu-239 Q Alarm Thresho	ld: et
Neutron Rate (CPS): Threshold: 500 Set	Pu-239 U-235 Ba-133 No Dwell Time (Second 1 2 3 4	eutron nds)  5 6
Back		Back



### **Advanced Setup Screens**









### **Advanced Setup Screens**



PocketPC     Iools Help	Eile Zoom Iools Help
Adv.         09/10/2011 12:02:54 AM           Setup         Battery Time: 164 min	<i>Adv.</i> 09/10/2011 12:02:09 AM <i>Setup</i> Battery Time: 165 min ————————————————————————————————————
Ask for Identifier Description	Data Location:
Default Description:	SD Card       ▼         Filename       ▲         2010_08_10_21_56_350.spc       ≡         2010_08_10_23_02_360.spc       □         2010_09_01_05_33_150.spc       □         2010_09_01_08_14_410.spc       □         2010_09_01_09_04_250.spc       □         2010_10_20_17_47_290.spc       □         2010_10_21_04_53_420.spc       ▼
Display Delete Back	Choose Back



## **Collecting a Spectrum**




PocketPC		
Ele Zom Tools Help SURVEY 09/09/2011 1 Mode Battery Tin Neutron Count Rate γ Dose Rate = 0.02 γ Count Rate = 2 Status: READY	ne: 165 min = 1 cps µSv/h	To collect a gamma spectrum, click on <b>Identify</b>
Search Display SNM Advanced	<b>I</b> dentify	



Cesium-137 (<sup>137</sup>Cs)




#### Three screens are available to monitor the data





### Logarithm vs Linear Y-Axis





Log scale provides "best" overall view of spectrum



### **Confidence Levels**



#### **Longer Collection Times = Higher Confidence**

PocketPC File Zoom Tools Help		<b>PocketPC</b> <u>File Zoom Tools H</u> elp			<b>PocketPC</b> <u>Fi</u> le <u>Z</u> oom <u>T</u> ools <u>H</u> el	þ	
ID Mode 🍯	Elapsed Time: 30 sec Battery Time: 165 min	ID Mode 🤇	Elapsed Tim Battery Time		ID Mode		ime: 301 sec pace: 12192
	Intense		– Intense –			Intense	
Nuclide CS137 Am241 W188 At211 Ba133 Se75 Lu172 Th232	Energy(keV)Confidence661.6055.95662.4555.95290.671.92669.601.73302.001.68279.501.511093.671.41911.001.41	Nuclide Am241 CS137 K40 I123 Th228 Th228 Ta182	Energy(keV) 0 662.45 661.60 1460.75 158.97 2614.53 583.20 1221.41	Confidence 112.08 112.07 3.46 2.47 2.00 1.84 1.73	Nuclide Am241 CS137 K40 Y88 Ga64 Sr-Y90 Pa233	Energy(keV) 662.45 661.60 1460.75 898.00 991.56 1761.00 300.34	Confidence 178.17 178.17 5.92 2.56 2.24 2.00 1.73
Save	Display Back	Save	Display	Back	Save	Display	Back

**Recommend collecting data for 300 seconds (5 minutes)** <sup>18</sup>



### **Radioisotope Library**



# Detective has a library *"look-up"* table with up to 80 radioisotopes and hundreds of gamma-rays

Am241 Cs137 Co60 Co57 Mn54 Th228 Th228 Eu152 K40 >NORM K40 Np237 Y88 TL201 TL202 TL202 TL204 Co56 SrRb82 U238	Cf252-249 Ba133 Shielded Ba133 Bi207 Ta182 Ho166 Ir194 Ir192 Po210 Positron emitter In111 La140(FP) Mo99 Mo + Tc99M Na22 Pu239	Sn113 Sr-Y90 Tc99M Ga67 Cu64 I123 I131 Xe131M Xe133 Xe135 Ra226 I125 At211 Br76 Cr51* Eu155* Eu156* Fe59*	Ga64* Gd153* Gd159* Pa233* Tl208* Ho166m* Pd103* Au198* Sm153* Lu177* Shielded Lu177* Lu172* W188* Re188* Se75* U233* U235

**Radioisotopes in library** 

	524.40	Pu-233	1.300000e-003	
	529.87	I-133	8.630000e-001	
	534.80	Pu-233	9.000000e-003	
	535.20	N-18	2.850000e-002	
	536.09	I-130	9.900000e-001	
	537.32	Ba-140	2.500000e-001	
	549.70	Rn-220	1.000000e-003	
	558.80	Pu-233	2.700000e-003	
	560.13	Pu-245	5.440000e-004	
	569.31	Cs-134	1.543000e-001	
	572.90	U-242	3.600000e-001	
_	583.10	TI-208	8.420000e-001	
	583.14	Th-232	3.083000e-001	
	583.30	Pu-233	8.600000e-004	
	585.00	U-242	3.700000e-001	
	600.56	Sb-125	1.780000e-001	
	602.71	Sb-124	9.787000e-001	


#### Look-up table is a spreadsheet with radioisotope data 19




## **Radioisotope ID**



#### How does a Detective identify a radioisotope?



Algorithm finds strongest peak(s) in spectrum



Energy is determined from calibration

524.40 529.87 534.80 535.20 536.09 537.32 549.70	Pu-233 I-133 Pu-233 N-18 I-130 Ba-140 Rn-220	1.300000e-003 8.630000e-001 9.000000e-003 2.850000e-002 9.900000e-001 2.500000e-001 1.000000e-003
558.80	Pu-233	2.700000e-003
560.13	Pu-245	5.440000e-004
569.31	Cs-134	1.543000e-001
572.90	U-242	3.600000e-001
583.10	TI-208	8.420000e-001
583.14	Th-232	3.083000e-001
583.30	Pu-233	8.600000e-004
585.00	U-242	3.700000e-001
600.56	Sb-125	1.780000e-001
602.71	Sb-124	9.787000e-001


Energy is "looked-up" in library to find radioisotope



### **Radioisotope ID**



# Why does it identify so many radioisotopes?



#### The look-up table has a large number of radioisotopes which have similar gamma energies


524.40	Pu-233	1.300000e-003
529.87	I-133	8.630000e-001
534.80	Pu-233	9.000000e-003
535.20	N-18	2.850000e-002
536.09	I-130	9.900000e-001
537.32	Ba-140	2.500000e-001
549.70	Rn-220	1.000000e-003
558.80	Pu-233	2.700000e-003
560.13	Pu-245	5.440000e-004
569.31	Cs-134	1.543000e-001
572.90	U-242	3.600000e-001
583.10	TI-208	8.420000e-001
583.14	Th-232	3.083000e-001
583.30	Pu-233	8.600000e-004
585.00	U-242	3.700000e-001
600.56	Sb-125	1.780000e-001
602.71	Sb-124	9.787000e-001



## **Radiation Background**



#### What does the radiation background look like?



The natural radiation background is complex containing minute quantities of radioactive isotopes of <sup>40</sup>K, <sup>238</sup>U and <sup>232</sup>Th and their decay products




### **Radiation Background**




# It can take 300 seconds or more to identify the background

Eile



PocketPC e <u>Z</u> oom <u>T</u> ools <u>H</u> elp			X
ID Mode 🥥		ime: 180 se me: 164 mii	
Nuclide	Energy(keV)	Confidence	
Th228 Th232 K40 Ra226 Ra226 Lu172	238.63 911.00 1460.75 351.87 609.00	4.24 3.61 3.58 3.32 3.19	
Cu64	PocketPC     File Zoom Icols Help		
Ir194	ID Mode 🥥	Elapsed Tin Storage Spa	
	Neutron	Count Rate :	= 1 cps
	Y Dose I	Rate = 0.02	µSv/h
Save	100 10 Marker: 2669		0 Cnts
	Zoom <	> ><	Lin
	Save	Pause	Back






### **Radiation Background**



#### **Quality Control**


A good quality control check of the Detective is to know the *gamma count rate per second (cps) at 1 meter* with no source present at your equipment staging location.

Before taking the Detective out for a measurement, verify the background gamma count rate.






### What is Dead Time?



# Keep dead time less than 5% by moving away from source







## **High Count Rate Warnings**




#### Three ways to tell if count rate is too high

Text Warning

High Dead Time

#### Flashing Red Banner





### **Dead Time Example**







A cargo container filled with NORM can produce a *radiation field* significantly above background but still safe to ship

These radiation fields can cause the *dead time* of the detector to be quite high requiring measurements to be taken from over 10 m away

High *dead time* can produce artificial peaks in the gamma spectrum resulting in mis-identification









#### **Lesson Learned - Embedded Source**




If a RIID is too close to the Detective during a measurement, then the <sup>137</sup>Cs calibration source in the RIID can be detected

Keep RIID at least 5 meters away from Detectives







Background

#### Suspect: Cs-137

Found: Cs-137





## **Radioisotope Examples**



The following section includes examples of Detective screen displays for some of the most common industrial radioisotopes:

Cesium-137	( <sup>137</sup> Cs)
Cobalt-60	( <sup>60</sup> Co)
Bismuth-207	( <sup>207</sup> Bi)
Sodium-22	( <sup>22</sup> Na)
Americium-241	( <sup>241</sup> Am)
Europium-152	( <sup>152</sup> Eu)
Barium-133	( <sup>133</sup> Ba)


Quick Reference (most common in red)				
Industrial	<sup>137</sup> Cs, <sup>60</sup> Co, <sup>192</sup> Ir, <sup>241</sup> Am, <sup>57</sup> Co, <sup>133</sup> Ba, <sup>78</sup> Se			
Medical	<sup>99m</sup> Tc, <sup>201</sup> TI, <sup>131</sup> I, <sup>19</sup> F, <sup>67</sup> Ga, <sup>111</sup> In, <sup>123</sup> I, <sup>133</sup> Xe			
Natural (NORM)	<sup>40</sup> K, <sup>232</sup> Th, <sup>238</sup> U			
Nuclear	<sup>235</sup> U, <sup>239</sup> Pu, <sup>237</sup> Np, <sup>233</sup> U			



Cesium-137 (<sup>137</sup>Cs)



### <sup>137</sup>Cs has a gamma-ray at 661.6 keV and is an industrial source






### **Cobalt-60 (60Co)**



### <sup>60</sup>Co has gamma-rays at 1173.2 and 1332.5 keV and is an industrial source






**Bismuth-207** (<sup>207</sup>Bi)



### <sup>207</sup>Bi has multiple gamma-rays and is an industrial source






Sodium-22 (<sup>22</sup>Na)



### <sup>22</sup>Na has gamma-rays at 511 and 1274.5 keV and is an industrial source






## Americium-241 (<sup>241</sup>Am)



### <sup>241</sup>Am has multiple gamma-rays and is an industrial source






**Europium-152 (<sup>152</sup>Eu)** 



### <sup>152</sup>Eu has multiple gamma-rays and is an industrial source

PocketPC



E	ile zoom Iools Help ID Mode	-	Elapsed Time: 60 sec Battery Time: 164 min Intense			
	Nuclide Co57 In111 Mo99 Eu152 Eu152 SR RB52 Au198 Ho166	Energy(keV) 122.06 245.35 777.90 1408.00 1112.07 776.52 411.80 410.90	Confidence 68.30 25.61 18.48 17.96 16.37 14.23 8.20 8.20			
	Save	Display	Back			

PocketPC Eile Zoom Tools Help						
IDElapsed Time: 43 secModeBattery Time: 164 min						
Neutror	Neutron Count Rate = 0 cps					
Y Dose	γ Dose Rate = 0.84 µSv/h					
	Found Industrial(1) Found Other(1)					
Intense	Display	Search				
Save	Save Pause					
		25				




Barium-133 (<sup>133</sup>Ba)



### <sup>133</sup>Ba has multiple gamma-rays and is an industrial source

PocketPC



Ele zoom Iools Help ID Mode 🌰	-	Elapsed Time: 61 sec Battery Time: 164 min		
Nuclide Ba133 Pd103 Xe133 Tl204 Ba133 Ba133 Ga67 Pa233 I123	<ul> <li>Intense</li> <li>Energy(keV)</li> <li>356.00</li> <li>357.45</li> <li>81.00</li> <li>80.20</li> <li>302.00</li> <li>383.00</li> <li>300.22</li> <li>300.34</li> <li>158.97</li> </ul>	Confidence 246.65 232.85 186.40 186.39 138.30 87.88 17.24 17.24 8.40		
Save	Display	Back		

	<b>PocketPC</b> <u>File Z</u> oom <u>T</u> ools <u>H</u> elp						
	<i>ID</i> Elapsed Time: 25 sec <i>Mode</i> G Storage Space: 12192						
	Neutron Count Rate = 0 cps						
	<del>y Duse Ra</del> te = 8.82 µSv/h						
(	Found Industrial(1)						
	Intense	Display	Search				
	Intense	Dispidy					
	Save	Pause	Back				



### **Training and Exercise**



Three HPGe Detectives will be setup with laptops, projectors and screens so that each group can view their PDA display

The instructor will divide the class into 3 groups and assign each group an expert to assist and answer questions

Each group will measure a series of radiation sources

For the Training Session, the radiation sources will be known

For the Exercise Session, the radiation sources will be unknown

All measurements will be saved and documented with a checklist



# **Spectroscopy Checklist**



	First Responder Gamma Spectroscopy Checklist - Group #						
Initial Setup Date/Time	Detective Serial Number	Detective Operating Status (ok)	Gamma Background Count Rate	Gamma Background Dose Rate	Background Measurement (300 seconds) Filename	Calibration Measurement (300 seconds) Filename	Calibration Source Activity/Date/Serial Number (kBq)
Source Measurements	Measurement (300 seconds) Filename	Detective Operating Status (ok)	Gamma Background Count Rate	Gamma Background Dose Rate	Measurement Dead Time (%)	Detection Distance (cm)	Radioisotopes Identified and Confidence Levels
Source 1							
Source 2							
Source 3							
Source 4							
Source 5							
Source 6							
Source 7							
Source 8							
Source 9							
Source 10							

38



# Training



**Review advanced setup screen and check operating status** 

Determine background count and dose rate as quality control check

Calibrate using <sup>137</sup>Cs and automated calibration feature

Collect and save 5 minute spectra for background and calibration

Check dead time for all measurements and adjust distance to source as necessary to obtain 5% or less

Collect and save 5 minute spectra for a series of known radiation sources

Complete a checklist for each radiation source

The training will be followed by questions and discussion



### Exercise



#### Scenario

Intelligence sources intercept documents which indicate that an abandoned warehouse near the border will be used as a transshipment point for a smuggled weapons cache. The date is close to an upcoming national event with international participants.

A law enforcement team supported by an Explosive Ordnance Disposal (EOD) unit raids the warehouse and using their radiation pagers, locate several containers of radioactive material. The radiation pagers provide readings ranging from 3 to 7 on contact.

The EOD unit requests an expert radiological team collect data on the containers to identify the radioactive material and provide an assessment of their results.



### Exercise



Determine background count and dose rate as quality control check

Calibrate using <sup>137</sup>Cs and automated calibration feature

Collect and save 5 minute spectra for background and calibration

Check dead time for all measurements and adjust distance to source as necessary to obtain 5% or less

Collect and save 5 minute spectra for a series of unknown radiation sources

Complete a checklist for each radiation source

Each group will provide a summary of their results and initial assessment

The exercise will be followed by questions and discussion





# **Questions?**

